

CATALOG

VeroCor

COPPER CORROSION BATH



VeroCor

TECHNICAL PARAMETERS

Solid block bath for silver and copper corrosion testing per ASTM D7667 & ASTM D130.


The solid block bath is specifically designed for measuring thin silver strip corrosion to ASTM D7667 and copper corrosion to ASTM D130.

FEATURES

- Two Groups of three sample positions, each group can work separately from the other group
- Aluminum Dry Block, with internal temperature monitoring sensors.
- Universal bath with special inserts to accommodate different sample cups depending on the sample and test standard.
- Fast heating, reaching the test temperature within 3 minutes.
- Maximum Temperature of 180 °C
- Safety measures to prevent overheating are provided through digitally controlled temperature settings secured by thermal semiconductor devices.

SPECIFICATIONS

| | |
|--------------------------|--|
| Heating System | Aluminum Dry Block |
| Sample Positions | Each group of three has its own independent and control system. |
| Power Consumption | Total power consumption is 1200 Watts. 600 Watts for each group. |
| Test Standards | Universal System to cover different test standards. |
| Power Requirement | 220 V, 50 to 60 Hz |
| Weight | 12 kg |



APPLICABLE

REFERENCES

REFERENCES



Vero
Scientific



VeroVis
Kinematic Viscometer



VeroRes
Carbon Residue
Analyzer



VeroXi
Oxidation Stability
Analyzer



VeroAsh
Ash Content Furnace



VeroDry
Moisture Evaporator



VeroCal
Heat Com. Calorimeter



VeroCor
Copper Corrosion Bath



VeroFoam
Foam Tendency
Analyzer

PHONE
+1781818 8133

EMAIL
sales@verosci.com

WEBSITE
www.verosci.com

ADDRESS
Vero Scientific LLC
867 Boylston Street
5th Floor #1409
Boston, MA 02116
United States

Copyright © 2025
Vero Scientific
All rights reserved.



As Vero Scientific, we provide advanced measurement technologies, preventive maintenance solutions, and cloud-based analytics for oil and fuel analysis. Our mission is to empower industries with precise, reliable, and sustainable solutions.